

# SIS RESULTS OF LAL TESTING

Sensitivity: 0.25 EU/ml

	Sample ID	No Dilution	1:4 dilution	1:10 dilution
Rubber only	# 1	+	-	-
"	# 2	+	+	+
Scrub2	# 3	-	NA	NA
Scrub2	# 4	-	NA	NA
CS-PAA 50°C-2h	# 5	-	NA	NA
"	# 6	-	NA	NA
20°C-2h	# 7	+	+	-
"	# 8	+	+	-
CS-PAA 40°C-2h	# 9	-	NA	NA
"	# 10	-	NA	NA

**Exhibit A**